



U.S. Corporate Headquarters
400 Valley Rd.
Warrington, PA 18976
1(800) 523-2575 / (215) 343-6484
1(800)343-3291 fax
info@polysciences.com

Polysciences Europe GmbH
Badener Str. 13
69493 Hirschberg an der Bergstrasse,
Germany
+(49) 06201-845200
+(49) 06201-8452020 fax
info@polysciences.de

Polysciences Asia-Pacific, Inc.
2F-1, 207 DunHua N. Rd.
Taipei, Taiwan 10595
(886) 2 8712 0600
(886) 2 8712 2677 fax
info@polysciences.tw

TECHNICAL DATA SHEET 555

Page 1 of 2

BioMag[®] Dextran-coated Charcoal

Catalog Number: 84510 & 84555

DESCRIPTION

BioMag[®] Dextran-coated Charcoal is a suspension of BioMag[®] particles covalently attached to NORIT[®] activated carbon and dextran. The suspension is supplied in distilled water with 0.1% sodium azide added as a preservative. A 1:5 dilution of this concentrate with phosphate buffered saline (PBS) is necessary before use. After shaking vigorously or vortexing, the diluted suspension is ready to use.

USE AND APPLICATIONS

BioMag[®] Dextran-coated Charcoal (5mg/mL preparation) is used in tritiated radioimmunoassays and other applications to simplify charcoal-based purifications. The dextran coating on the charcoal serves a gatekeeping or sieving function, permitting small molecules to pass between dextran molecules into the charcoal absorbent, while excluding larger complexes. Free unlabeled and radiolabeled analyte may thus be separated from the antibody-bound fraction. The dextran coated particles have also been used for fluorescent-based assay cleanup, most notably in rapid transglutaminase assays for high-throughput screening applications which employed dansylcadaverine. Similar peptide-based fluorescent assays may also benefit from this cleanup strategy as well, or other small fluorescent molecule labeling, although it's recommended the molecular weight of the binding biomolecule partner be in excess of 10 kD to limit loss.

CHARACTERISTICS

Mean Diameter: 1-10 μ m
Particle Concentration: 84510, 5mg/mL, 84555, 50 mg/mL
Binding Capacity: 750 μ L (3.75mg) of dilution per assay tube is sufficient to adsorb free unlabeled and tritium (³H)-labeled analyte in a typical competitive radioimmunoassay. *Use only the diluted product in immunoassays.*

MATERIAL

Material Supplied

BM555 BioMag[®] Dextran-coated Charcoal: 100mL; or BM556 BioMag[®] Dextran-coated Charcoal Concentrate: 1000mL

Material Required

Phosphate buffered saline (if needed for dilution)
Magnetic separator
Test tubes
Scintillation vials

PROCEDURE

Researchers are advised to optimize the use of BioMag[®] in any application as procedures designed by other manufacturers may not be ideal.

In preparation for use, the product should be shaken vigorously as well as before dilution. If working with the concentrated form (BM 556, 50mg/mL), a 1:5 dilution of this concentrate with phosphate buffered saline (PBS) is necessary before use. Add 4 parts PBS to 1 part concentrate.

Assay tubes are placed in a magnetic separation unit that pulls BioMag[®] Dextran-coated Charcoal Concentrate containing the free radiolabeled and unlabeled analyte to the bottom of the test tubes, leaving the antibody bound analyte in the supernatant. The entire separation unit is inverted to simultaneously decant thirty test tubes into scintillation vials to quantify antibody-bound analyte. Please inquire for further information on the BioMag[®] separation device best suited to your application.

REFERENCES

- Herbert, V., K.S. Lau, C.W. Gottlieb, S.J. Bleicher. 1965. *Coated charcoal immunoassay of insulin*. J Clin Endocrinol Metab, 25(10): 1375-1384.
- Li, H., Zhang, L., Cui, Y., Luo, X., Xue, C., Wang, S., ... & Du, M. (2014). *Characterization of recombinant Zea mays transglutaminase expressed in Pichia pastoris and its impact on full and non-fat yoghurts*. Journal of the Science of Food and Agriculture, 94(6), 1225-1230.
- Naik, A. D., Menegatti, S., Reese, H. R., Gurgel, P. V., & Carbonell, R. G. (2012). *Process for purification of monoclonal antibody expressed in transgenic Lemna plant extract using dextran-coated charcoal and hexamer peptide affinity resin*. Journal of Chromatography A, 1260, 61-66.
- Poznanski, N., W.J. Poznanski. 1969. *Laboratory application of the dextran-coated-charcoal radioimmunoassay of insulin*. Clin Chem, 15(9):908-918.
- Wu, Y.W., Y.H. Tsai. 2006. *A rapid transglutaminase assay for high-throughput screening applications*. J Biomol Screening, 11(7): 836-843.
- Wu, Y. W., & Tsai, Y. H. (2012). U.S. Patent No. 8,133,683. Washington, DC: U.S. Patent and Trademark Office.

TRADEMARKS

BioMag[®] is a registered trademark of Polysciences, Inc.
NORIT[®] is a registered trademark of Norit N.V.

STORAGE AND STABILITY

Store at 2-8°C. Freezing, drying, or centrifuging particles may result in irreversible aggregation, with loss of surface area and binding capacity.

SAFETY

This particle suspension contains sodium azide. Sodium azide may react with lead and copper plumbing to form explosive metal azides. Upon disposal of material, flush with a large volume of water to prevent azide accumulation. Please consult the Safety Data Sheet for more information.

Observe your institution's guidelines for safe handling and disposal of radiolabeled materials if using BioMag® Dextran-coated Charcoal in radiometric assays or isolations.

This product is for research use only and is not intended for use in humans or for *in vitro* diagnostic use.

ORDERING INFORMATION

Cat. #	Description	Size
84510	BioMag® Dextran-coated Charcoal	100ml
84555-1000	BioMag® Dextran-coated Charcoal Concentrate	1000ml

TO ORDER

In The U.S. Call: 1(800) 523-2575 • (215) 343-6484
 In The U.S. Fax: 1(800) 343-3291 • (215) 343-0214
 In Germany Call: +(49) 06201-845200
 In Germany Fax: +(49) 06201-8452020
 In Asia Call: (886) 2 8712 0600
 In Asia Fax: (886) 2 8712 2677

Order online anytime at www.polysciences.com

Should any of our materials fail to perform to our specifications, we will be pleased to provide replacements or return the purchase price. We solicit your inquiries concerning all needs for life sciences work. The information given in this bulletin is to the best of our knowledge accurate, but no warranty is expressed or implied. It is the user's responsibility to determine the suitability for their own use of the products described herein, and since conditions of use are beyond our control, we disclaim all liability with respect to the use of any material supplied by us. Nothing contained herein shall be construed as a recommendation to use any product or to practice any process in violation of any law or any government regulation.